


Medical Devices Made With Polyvinylchloride (PVC) Using the Plasticizer di-(2- Ethylhexyl)phthalate (DEHP); Draft Guidance for Industry and FDA

Draft Guidance – Not for Implementation

**This guidance document is being distributed for comment purposes only.
Draft released for comment on September 6, 2002**



**U.S. Department Of Health and Human Services
Food and Drug Administration
Center for Devices and Radiological Health**

Office of Device Evaluation

Preface

Public Comment

Comments and suggestions regarding this draft document should be submitted by December 5, 2002 to Docket No. 02D-0325, Dockets Management Branch, Division of Management Systems and Policy, Office of Human Resources and Management Services, Food and Drug Administration, 5630 Fishers Lane, Room 1061, (HFA-305), Rockville, MD 20852.

Additional Copies

Additional copies are available from the Internet at:

<http://www.fda.gov/cdrh/ode/guidance/1407.pdf>, or CDRH Facts-On-Demand. In order to receive this document via your fax machine; call the CDRH Facts-On-Demand system at 800-899-0381 or 301-827-0111 from a touch-tone telephone. Press 1 to enter the system. At the second voice prompt, press 1 to order a document. Enter the document number (**1407**) followed by the pound sign (#). Follow the remaining voice prompts to complete your request.

1 **Medical Devices Made With**
2 **Polyvinylchloride (PVC) Using the**
3 **Plasticizer di-(2-Ethylhexyl)phthalate**
4 **(DEHP); Draft Guidance for Industry and**
5 **FDA**

6
7 *This document is intended to provide guidance. It represents the Agency’s current*
8 *thinking on this topic. It does not create or confer any rights for or on any person and*
9 *does not operate to bind the Food and Drug Administration (FDA) or the public. An*
10 *alternative approach may be used if such approach satisfies the requirements of the*
11 *applicable statute and regulations.*

12
13 **Who should read this guidance document?**

14
15 Manufacturers who fabricate their devices with polyvinylchloride (PVC) using the
16 plasticizer di-(2-ethylhexyl)phthalate (DEHP) should read this guidance.

17
18 **What is the concern that this document is attempting to address?**

19
20 DEHP is recognized as an important chemical ingredient that affords PVC many of the
21 physical properties that make the material optimally suited for use in many of today’s medical
22 devices. DEHP is a chemical whose long-term effects on the human body are unknown.
23 Through this guidance, FDA is offering suggestions on ways that you may reduce or eliminate
24 risks that may be associated with DEHP. We are suggesting that you label certain devices with
25 their DEHP content and consider eliminating the use of DEHP in certain devices that can result
26 in high aggregate exposures in sensitive patient populations.

27
28 **Does DEHP have adverse effects in humans?**

29
30 Although the toxic and carcinogenic effects of DEHP have been demonstrated in
31 laboratory animals, there are no human studies that show such effects. What we do know is
32 that there are certain invasive medical procedures during which exposure to DEHP could
33 exceed the levels that are not expected to cause any adverse health effects in patients.

34 **Does FDA have concern with all devices that contain PVC?**

35
36 No. As stated above, not all devices made with PVC contain DEHP. Further, FDA
37 recognizes that many devices with PVC containing DEHP are not used in ways that result in
38 significant human exposure to the chemical. Therefore, FDA is focusing attention on the small
39 subset of medical devices where PVC containing DEHP may come in contact with the tissue of
40 a sensitive patient population in a manner and for a period of time that may raise concerns about
41 the aggregate exposure to DEHP. We believe that many devices used in Neonatal Intensive
42 Care Units (NICUs) meet this criteria and should be a primary focus.

43
44 **What types of CDRH-regulated devices typically may be of concern?**

45
46 Again, the risks from DEHP-containing devices relate to: (1) the aggregate exposure
47 and (2) the sensitivity of the exposed patient population. While devices used in neonates
48 deserve particular attention, there may be other patient subgroups where DEHP exposure may
49 be an issue. The following types of medical devices may contain PVC components, e.g., tubing
50 or fluid containers, that could contain the chemical DEHP and expose sensitive patient
51 populations:

52
53 **Intravascular (IV) tubing and catheters/cannulae** used in:

- 54 → IV administration
- 55 → dialysis
- 56 → extracorporeal membrane oxygenation (ECMO)
- 57 → cardio-pulmonary bypass (CPB) procedures.

58 **Bags** used to store and transport:

- 59 → enteral nutrition formulae
- 60 → total parenteral nutrition formulae.

61 **Tubing** used in enteral nutrition:

- 62 → nasogastric tubes
- 63 → gastrostomy tubes
- 64 → nasojejunal tubes

65 **What does FDA recommend that you do if your device is made with PVC containing**
66 **DEHP?**

67
68 We encourage you to consider all mechanisms to reduce patient exposure to DEHP,
69 particularly from those types of devices cited above. Specifically, we recommend that you
70 consider the feasibility of replacing PVC containing DEHP with either alternative materials or
71 plasticizers, or using coatings that may minimize patient exposure to DEHP. Manufacturers
72 should consider “minimizing patient exposure to DEHP” as a design requirement in their design
73 control procedures (Quality System regulation, [21 CFR 820.30](#)). This step should contribute to
74 an overall reduction in patient exposure to DEHP.

75
76 **Do I need to submit a new 510(k) if I replace or modify the PVC in my device?**

77
78 Generally, 510(k) holders do not need to submit 510(k)s to make materials changes of the
79 nature discussed in this document. Nevertheless, you should refer to [21 CFR 807.81\(a\)\(3\)](#) and our
80 guidance document entitled, “*Deciding When to Submit a 510(k) for a Change to an Existing*
81 *Device*” (<http://www.fda.gov/cdrh/ode/510kmod.html>), to determine whether any particular change
82 requires the submission of a 510(k). It is your responsibility to determine whether you need to
83 submit a 510(k) before making any particular change and to document your decision-making
84 process in your design validation records.

85
86 **What should I do if my device is exempt from 510(k) and I replace or modify the PVC in**
87 **my device?**

88
89 If your class I or class II device is exempt from the 510(k) requirements of the Federal
90 Food, Drug, and Cosmetic Act, you should refer to the limitations of exemptions from section
91 510(k) to determine whether any particular change exceeds the limitations of the exemption,
92 thereby requiring the submission of a 510(k). Generally, it is unlikely that materials changes of
93 the nature discussed in this document would exceed the limitations of exemptions. The
94 limitations of exemptions are codified at 21 CFR XXX.9, where XXX is the part of title 21 of
95 the Code of Federal Regulations containing the particular device’s classification, (e.g., [21 CFR](#)
96 [880.9](#) for Intravascular Administration Sets [21 CFR 880.5440](#)).

97
98 **Do I need to submit a PMA Supplement if I replace or modify the PVC in my class III**
99 **device?**

100
101 Generally, manufacturers of class III devices must assess each individual change to
102 determine whether the change affects the safety and effectiveness of the device ([21 CFR 814.39](#)),
103 thereby requiring the submission of a PMA supplement. However, FDA believes most
104 manufacturers will not need to submit PMA supplements to make materials changes of the nature

105 discussed in this document. Manufacturers may conclude that such changes do not require prior
106 FDA approval when the change is to a material known to be suitable for the particular use, and
107 when verification and validation studies confirm that the device use requirements continue to be met.
108 In cases where PMA supplements are not required, manufacturers should provide notification of the
109 particular change in the next annual report to the PMA.

110

111 **What if I choose not to change the material in my device? Should I revise the labeling to**
112 **state the device contains DEHP?**

113

114 Yes, we recommend that you clearly indicate through user labeling that your device contains
115 DEHP. Although, at this time, FDA believes there is insufficient information to justify requiring
116 device manufacturers to disclose the presence of this chemical in the device's labeling, there is
117 considerable interest among some consumers and practitioners in mitigating any risks that exposure
118 to DEHP may present. Disclosure can assist healthcare professionals in making informed decisions
119 regarding an individual patient's exposure to DEHP.

120

121 **Do I need to submit a new 510(k) or PMA Supplement if I revise the labeling to state**
122 **the device contains DEHP?**

123

124 This change does not require a new 510(k) or PMA Supplement. However, you should
125 document the change in your device master record. For PMA devices, you should submit the
126 change in your next Annual Report to the PMA.

127

128 **Where can I find more information about medical devices and DEHP?**

129

130 The Center for Devices and Radiological Health (CDRH) recently released a safety
131 assessment, which can be found at <http://www.fda.gov/cdrh/ost/dehp-pvc.pdf>. In this document,
132 the doses of DEHP received by patients undergoing various medical procedures are compared to
133 the Tolerable Intake (TI) values for DEHP using ISO/DIS 10993-17 *Method for the*
134 *Establishment of Allowable Limits for Leachable Substances*.